GOVT. OF NCT OF DELHI
AMBEDKAR NAGAR HOSPITAL,
BLOCK-B, SECTOR-5, DAKSHINPURI, NEW DELHI-110062
Email ID: mdanhospital@gmail.com

Date: 13/10/2026

F.3(88)/Ophthalmology/Pur./ANH/2026/ 6948 - 6950

NOTICE INVITING QUOTATION

Price Quotations are invited from interested Vendors/Dealers etc. for procurement of following items for Ophthalmology Department at Ambedkar Nagar Hospital, Dakshinpuri. Only sealed quotations for the required items mentioned in the table below may be submitted by interested Vendors/Dealers etc. latest by 27.10.2026, up to 4 pm at Ambedkar Nagar Hospital, Dakshinpuri, Delhi-110062.

Sl. No.	Item Name	Department	Quantity/ No.	Unit price	GST in %	Total price with GST
1.	Direct Ophthalmoscope with streak retinoscope set	Ophthalmology	01			
2.	90 D lens		01			
3.	Goldman 3 mirror gonioscopy lens		01			

Note: All the Vendors/Dealers should submit the following undertaking along with their quotations.

"I hereby undertake that; I am reliable and competent to supply the services in question and that I am/my firm is not debarred or blacklisted by Department of Commerce or Ministry/Department concerned."

Other terms & conditions: -

1.Quotation No and Name must be written on top of the envelope.
2.Rate should be mentioned in word and figures both in the quotation submitted.
3. GST (Kindly mention in above table) should be clearly mentioned in the offer.
4. Delivery Schedule - within 15 days from the date of issue of Purchase Order.
5.LD @ 0.5% of delayed supply per week or part of week for delay of supply of material subject to maximum up to 10% of delayed supply should be deducted.
6. Validity of offer should not be less than 90 days.
7.Firm should provide **onsite Warranty** and CMC for period as per attached specifications of concerned equipment.
8.100% payment will be made against satisfactory item verification by the concerned user department.
9.Item wise evaluation of the required equipment mentioned in the NIQ will be done and L1 quoter firm shall be selected accordingly for each item in the quotations submitted by the firms.

MoI/C (Purchase)

F.3(88)/Ophthalmology/Pur./ANH/2026/ 6948 - 6950

Date: 13/10/2026

Copy to:-

i.All MS/MD of hospitals of GNCTD through PA to MD, ANH.

ii.MoI/c website for uploading on official website of the Hospital.

14/10/26 DA (website)

iii. Notice Board.

MoI/C (Purchase)

SPECIFICATIONS OF DIRECT OPHTHALMOSCOPE WITH STREAK RETINOSCOPE

- Handheld direct ophthalmoscope with dust free sealed optics with rubber brow rest
- LED lamp illumination
- Rechargeable handle (3.5V), stepless dimming with charge status indicator with Li-ion rechargeable battery with charger
- Bayonet fitting for fast and secure attachment to ophthalmoscope and retinoscope head
- Glass–fibre reinforced casing which is sturdy, light weight and durable
- Simple exchange of lamp at base of instrument head
- Power range of lenses: Plus: +1 to +38, Minus: -1 to -30
- Easy to operate aperture hand wheel with aperture-filter combo:
  - Slit, fixation star with grade , semi circle, small and large aperture
- Filter wheel engageable for all apertures: red free (green), cobalt blue, polarization filter
- Focusing device with zero retention force for quick correction of ametropias
- Streak retinoscope head of same company compatible with same handle as ophthalmoscope
  - Simple operation with knurled thumb screw
  - Slit image can be focused with operating element and turned 360°
  - Holder for hanging and fixing the fixation cards
  - Simple exchange of lamp at base of instrument head
- Set to be provided with
  - Original manufacturer case
  - Original rechargeable Li-ion battery : 1
  - Fixation cards : 2-3
  - Original spare bulbs : 10
- Warranty : 2 years + 8yr cmc
- Certification : BIS/ISO/US FDA/CE
- Other tender conditions:
  - Suppliers should have been in the market for at least 3 years
  - Suppliers should have made a large number of installations within the last five years in the country in reputed institutions preferably in government hospitals with a proven track record of excellent after sales support for this system

Dr. Kavitha Anand
Specialist
Ophthalmology
AIIH

SPECIFICATIONS OF 90D LENS

- Double aspheric lens, superior quality
- Anti reflective coating
- Image magnification: 0.75X
- Laser spot magnification: 1.32X
- Field of view: 74° / 89°
- Working distance: 7mm
- Warranty : 2 yr

Dr Karthika
Consultant
Specialist
Ophthalmology
AIMS

SPECIFICATIONS OF GOLDMAN GONIOSCOPY 3-MIRROR LENS WITH FLANGE

- Mirror angle: 60 °/ 66 ° / 76 ° mirror angle
- Image magnification: 1.06x
- Laser spot size: 0.94 X
- Contact diameter: 15mm
- Warranty : 2yr

Dr. Kavitha
Specialist
Ophthalmology
AIMH